



A pragmatic approach to AI in HealthTech

A focus on empowering the Quality Assurance and Regulatory Affairs (QA/RA) professional

26 Aug 2025



Disclaimer

The analyses, their interpretation, and related information contained herein are made and provided subject to the assumptions, methodologies, caveats, and variables described in this report and are based on third party sources and data reasonably believed to be reliable. No warranty is made as to the completeness or accuracy of such third party sources or data.

As with any attempt to estimate future events, the forecasts, projections, conclusions, and other information included herein are subject to certain risks and uncertainties, and are not to be considered guarantees of any particular outcome.

All reproduction rights, quotations, broadcasting, publications reserved. The contents of this presentation are confidential, and no part of this presentation may be reproduced or transmitted in any form or by any means, electronic or mechanical, including photocopy, recording, or any information storage and retrieval system, without express written consent of IQVIA.

Copyright © 2025 IQVIA. All rights reserved. IQVIA® is a registered trademark of IQVIA Inc. in the United States and various other countries.

Today's presenters



Mike King
*Senior Director,
Strategy and Commercialization*

Mike is the Senior Director of Strategy and Commercialization within the Technology Solutions business of IQVIA.

- ❖ Responsible for optimizing business workflow across the Regulatory, Quality and Safety Solutions.
- ❖ 20+ years leading global teams in Regulatory Affairs and Quality Assurance across a range of therapeutic areas

Michael.King@iqvia.com



Anusha Gangadhara
*Associate Director,
Product Owner*

Anusha is the RIM Smart Product Manager and MedTech SME.

- ❖ Responsible for product definitions built from business requirements
- ❖ Define functionality of solutions to meet Global Regulations.
- ❖ 15+ years leading global teams in Regulatory Affairs and Quality Assurance.

Anusha.Gangadhara@iqvia.com



Todd Neal
*Senior Product Manager,
IT Design & Development*

Todd is the Senior Product Manager for a range of end-to-end processes in IQVIA's Quality Management System solution

- ❖ Responsible for Complaint Management, Post-Market Surveillance and Supplier Management modules of
- ❖ 30+ years experience in Enterprise Software and in developing solutions for healthcare

Todd.Neal@iqvia.com

To what extent is AI used within your company QMS processes?

Please select the most suitable below:

1. No AI is used
2. We are considering the use of AI in some targeted QMS processes
3. Some of our QMS processes use AI, however we are looking to broaden its use
4. AI is used to a large extent, within appropriate processes and where suitable for the organization.

Navigating the world of QA/RA in the era of AI enabled solutions

Mike King
Senior Director, Strategy & Commercialization




Global healthcare markets are progressively challenging

Use targeted AI solutions to drive a dual focus on patient safety and commercial performance

The problem statement:


MedTech and Pharma industries strive for global market access frameworks that are:

- 1 Aligned to commercial and strategic objectives
- 2 Efficient, timely and repeatable
- 3 Effective in adapting to global compliance process variations



Artificial Intelligence (AI) is a broad set of complementary methods

Be aware of the risk-benefit(s) of the different types of AI that are available



Global AI regulations continue to advance

Maintain oversight of horizontal AI regulations alongside vertical healthcare regulations and standards



EU AI Act: The European Union's AI Act classifies AI applications by risk level, with stringent requirements for high-risk applications such as biometric identification and financial decisions. It emphasizes data governance, transparency, and human oversight.



US AI Guidelines: In the United States, the National Institute of Standards and Technology (NIST) has developed the AI Risk Management Framework, which provides guidelines for managing risks associated with AI systems, particularly in critical sectors like healthcare and finance.



UK AI Regulation: The UK has adopted a pro-innovation approach, allowing individual regulatory bodies to govern AI within their domains. This includes guidelines from the Financial Conduct Authority (FCA), the Information Commissioner's Office (ICO) along with MHRA guidelines. This cross-industry approach is particularly useful as it enables best practices to be shared across industries.



Japan: Japan has introduced the Bill on the Promotion of Research, Development and Utilization of Artificial Intelligence-Related Technologies (AI Bill), which was submitted to Parliament on February 28, 2025. This bill aims to position Japan as a leader in AI innovation by adopting a lighter regulatory approach compared to other countries.




South Korea: Introduced the AI Basic Act, which is set to take effect in January 2026. This act consolidates 19 separate AI-related regulatory proposals into a unified framework, making it the second comprehensive AI-specific legislation in the world, following the EU's AI Act.



China: Has implemented several AI-related regulations, focusing on data security, privacy, and ethical use of AI. The country has also published guidelines for AI development and deployment.

Navigate complexity by solutioning out from targeted use cases

Be specific and focused – a default fallback is legacy solutions and/ or personal herculean effort



Relevant

- What is my **QA/RA use case(s)**?
- What **process steps** are outputs are defined by regulations and standards?



Controlled

- What **healthcare verification & validation** requirements are required? (e.g. US 21 CFR)
- What other **regulations** apply? (e.g. EU AI Act, EU GDPR, Cyber security,...)



Cost Effective

- What **solutions** are available, inclusive of AI?
- What is the **risk benefit / value** of each?
- What is the **cost** to implement and maintain?



Empowering QA/RA professionals with AI enabled solutions drives clear value



Patient Outcomes



Product Quality



Commercial Performance

Agentic AI: Transforming the role of QA/RA

Anusha Gangadhara
Associate Director, Product Owner



Harvesting the full product lifecycle – to power future workflow

Traceability and precedence through strong connected data frameworks across QA/RA systems



- Which countries should be targeted **first for the global launch** activities, which ones next?
- What is the **optimal regulatory pathway** enabling me to get to market faster?
- How can I optimize my lifecycle maintenance activities to **maximize re-use and reduce workload** of HA submissions?
- What procedure in which country? what is the **submission timeline**? what is the **submission format, fee**?, what pre-conditions apply? What documents can be reused?
- What **documents must be submitted**? must the document be legalised?

Harvesting the full product lifecycle – to power future workflow

AI Insights, impact assessment and predictive analysis need to stitch the full ecosystem



Built and tracked on Global and in-country Standards, Regulations and Governance

Need for validated 'Dynamic Data' throughout QA/RA processes

Transformation from manual to automated intelligence

Global lists

- Change details to global documents + ISO/IEC
- Standards to product characteristics
- Product Characteristics + ISO/IEC
- Complaint codes (IMDRF) to global documents

Country lists

- Agency (notified bodies included) & submission types
- Language, climatic zone
- Country to risk classification per regulation
- Submission templates – 50+ countries for Medical Devices
- License validity period per country
- Reference markets
- PMS activities list
- Initial submission fee + timeline
- Legal representative/sponsor requirement

IMDRF Global STED mapping


- STED-EU-FDA
- Automapping to country Templates

Process Examples

- **Impact Assessment to identify affected**
 - Products
 - Processes
 - In-country registrations
 - Global documents
- **Global registration planning**
 - Driven by impact assessment
 - Triggered by change management, standards or regulation changes and complaints
 - Insights on reference markets and delta submission requirements for market expansion
- **Drive In-country registration and submission build**
 - In-country registration planning with average timelines and fee estimation
 - Insights from precedence data on agency interactions and affected areas
 - Track and plan renewals and expiry
 - Use global submission documents to drive population of regional content

Building AI features as core platform capabilities

Enhancing QA/RA's focus on patient safety, product quality and commercial performance




- Gen AI prompts
- NLP (100+ languages)
- Similar event recognitions using precedence
- Intelligent summarization
- Intelligent recommendation assistant
- Decision tree-based recommendation
- Text and Image ingestion


Using AI in defining and building dynamic QA/RA workflows


Define the problem statement & build the workflow using regulatory intelligence and precedence

New draft global plan

Optimizing for shortest net timeline

 USA - FDA (primary market)


 EU - MDR

 Japan - PMDA / MHLW

+ Add

Clear

Generate Plan



Using AI in defining and building dynamic QA/RA workflows

Optimize precedence and recommendations to accommodate current scenarios

New draft global plan

Optimizing for shortest net timeline

USA - FDA (primary market)

EU - MDR

Japan - PMDA / MHLW

+ Add

Clear Generate Plan


Submission Timelines Reset to recommended | Region | Status | Regulatory Path | Preparation | Review | Total Timeline | |-------------------|--------|--------------------|-------------------------------------|--------------------|----------------| | USA - FDA | ✓ | 510(k) PMA De Novo | 6 months 12 months
Preparation | 177 days
Review | ~18 months | | EU - MDR | ✓ | CE Marking | - | - | ~1 year | | Japan - PMDA/MHLW | ✓ | Ninsho (PMC) | - | - | ~1 year |


Using AI in defining and building dynamic QA/RA workflows


Insights and impact assessments to define tasks and actions

New draft global plan

Optimizing for shortest net timeline

 USA - FDA (primary market)



 EU - MDR

 Japan - PMDA / MHLW

+ Add

Clear

Generate Plan



SmartSolve® eQMS Demonstration

Todd Neal
Senior Product Manager, IT Design & Development



Complaint Handling

Drive improved quality, volume and consistency whilst accelerating timeliness

Case Intake

- Utilising **NLP AI** to automate case intake
- Supports a variety of data sources

Adverse Event Reports

- Utilising **Gen AI** to improve in US MDR coding
- Trained using 5 years of US MAUDE data

Language

- Embedded **translation capabilities**
- Enter and review information in different languages



★ CASE-2025-0002

Case Intake

Case Coordinator: TNEAL Country of Origin: --- Occurrence Date: 23-Jan-2025 Case Type: TBD Case Severity: --- Reporting Site: NASH
Reportable?: TBD Case Aging Days: 211 Status: Inworks

Description as Reported *

During an in-clinic interrogation, an episode of over-sensed noise was observed. Technical Services were contacted and confirmed that poor telemetry might be the cause. The patient was scheduled for a follow-up visit the next day, but no additional details were provided. The device was recently removed and replaced to address the issue, and no further adverse effects were reported. The device is expected to be returned for analysis, although it has not yet been received.

▼ Contact Information (1)

▼ Incident Section

▼ Client Information

▲ Product Information (1)

+ Add

| Product | Complaint Number | Product Event Type | Aware Date |
|---------|------------------|--------------------|-------------|
| | | TBD | 23-Jan-2025 |

▼ Patient Information (0)

▼ Policy

▼ Due Date (# of business days)

▼ Team

▼ Attachments (0)

Cancel Save Sign-off

Product Information (1 of 1)

Manufacturing Site

Manufactured Date

DD-MMM-YYYY

Expiration Date

DD-MMM-YYYY

Expected Return Date

DD-MMM-YYYY

Complaint Quantity

1.00

Expected Return Quantity

0.00

No Product Return Rationale

▲ Product Failure Mode (1)

+ Add

| Component | Failure Mode | IMDRF Medical Device Problem Codes (Annex A) | Root Cause | Root Cause Category |
|----------------------|----------------------|--|----------------------|---------------------|
| <input type="text"/> | <input type="text"/> | <input type="text"/> | <input type="text"/> | Select Item... |

▼ External Reference (0)

IMDRF Dictionary



Search By

☒ Code/Term ☐ Definition ☐ All

Search Options

☒ All Words ☐ Either Word ☐ Exact Phrase

Search

[Display All](#)

▼ Select an IMDRF Annex A from the list below (IMDRF Code)

▼ Activation, Positioning or Separation Problem (A15)

▼ Activation Problem (A1501)

- Activation Failure (A150101)
- Difficult or Delayed Activation (A150102)
- Key or Button Unresponsive/not Working (A150105)
- Premature Activation (A150103)
- Self-Activation or Keying (A150104)
- Unexpected Deactivation (A150106)

▼ Positioning Problem (A1502)

- Difficult or Delayed Positioning (A150203)
- Difficult to Advance (A150205)
- Difficult to Insert (A150206)

Cancel

Run Assistant

Select

IMDRF Dictionary

Search By

☒ Code/Term ☐ Definition ☐ All

Search Options

☒ All Words ☐ Either Word ☐ Exact Phrase

Search

Search

Display All

▼ Select an IMDRF Annex A from the list below (IMDRF Code)

▼ Activation, Positioning or Separation Problem (A15)

▼ Activation Problem (A1501)

- Activation Failure (A150101)
- Difficult or Delayed Activation (A150102)
- Key or Button Unresponsive/not Working (A150105)
- Premature Activation (A150103)
- Self-Activation or Keying (A150104)
- Unexpected Deactivation (A150106)

▼ Positioning Problem (A1502)

- Difficult or Delayed Positioning (A150203)
- Difficult to Advance (A150205)
- Difficult to Insert (A150206)

Device Problem Code Assistant

| Term | IMDRF Code | FDA Code | Confidence % | Confidence Level |
|-----------------------|------------|----------|--------------|------------------------|
| Over-Sensing | A070909 | 1438 | 93.40% | <div><div></div></div> |
| Signal Artifact/Noise | A090801 | 1036 | 38.63% | <div><div></div></div> |
| Interrogation Problem | A0711 | 4017 | 12.63% | <div><div></div></div> |
| Pacing Problem | A0712 | 1439 | 10.92% | <div><div></div></div> |
| Telemetry Discrepancy | A1304 | 1629 | 4.48% | <div><div></div></div> |

Rows 5

1-5 of 10

< < Page 1 of 2 >

Cancel

Run Assistant

Select

IMDRF Dictionary



Search By ☒ Code/Term ☐ Definition ☐ All

Search Options ☒ All Words ☐ Either Word ☐ Exact Phrase

Search [Display All](#)

▼ Select an IMDRF Annex A from the list below (IMDRF Code)

▼ Electrical /Electronic Property Problem (A07)

▼ Device Sensing Problem (A0709)

Over-Sensing (A070909)



Problem related to failure of the device to properly filter cardiac signals resulting in inappropriate device response.

Device Problem Code Assistant

| Term | IMDRF Code | FDA Code | Confidence % | Confidence Level |
|---------------------------------------|------------|----------|--------------|------------------------|
| Over-Sensing | A070909 | 1438 | 93.40% | <div><div></div></div> |
| Signal Artifact/Noise | A090801 | 1036 | 38.63% | <div><div></div></div> |
| Interrogation Problem | A0711 | 4017 | 12.63% | <div><div></div></div> |
| Pacing Problem | A0712 | 1439 | 10.92% | <div><div></div></div> |
| Telemetry Discrepancy | A1304 | 1629 | 4.48% | <div><div></div></div> |

Rows

1-5 of 10

◀ < Page of > ▶

[Cancel](#)

[Run Assistant](#)

[Select](#)

★ CASE-2025-0002

Case Intake

Case Coordinator: TNEAL Country of Origin: --- Occurrence Date: 23-Jan-2025 Case Type: TBD Case Severity: --- Reporting Site: NASH
Reportable?: TBD Case Aging Days: 211 Status: Inworks

Description as Reported *

During an in-clinic interrogation, an episode of over-sensed noise was observed. Technical Services were contacted and confirmed that poor telemetry might be the cause. The patient was scheduled for a follow-up visit the next day, but no additional details were provided. The device was recently removed and replaced to address the issue, and no further adverse effects were reported. The device is expected to be returned for analysis, although it has not yet been received.

- ▼ Contact Information (1)
- ▼ Incident Section
- ▼ Client Information
- ▲ Product Information (1)

| Product | Complaint Number | Product Event Type | Aware Date |
|---------|------------------|--------------------|-------------|
| | | TBD | 23-Jan-2025 |

- ▼ Patient Information (0)
- ▼ Policy
- ▼ Due Date (# of business days)
- ▼ Team
- ▼ Attachments (0)

Cancel

Save

Sign-off

Product Information (1 of 1)

✕

Manufacturing Site

Manufactured Date

Expiration Date

Expected Return Date

Complaint Quantity

Expected Return Quantity

No Product Return Rationale

▲ Product Failure Mode (1)

+ Add

| Component | Failure Mode | IMDRF Medical Device Problem Codes (Annex A) | Root Cause | Root Cause Category |
|-----------|--------------|--|--------------|---------------------|
| | | A070909 | Over-Sensing | Select Item... |

▼ External Reference (0)

★ CASE-2024-0020-1

Product: DP Country of Origin: UNITED STATES Reportable?: Yes Complaint Coordinator: TNEAL Reporting Site: NASH Aware Date: 11-Dec-2024 Complaint Aging Days: 254 Case Severity: Serious Injury Status: Inworks

Task

Action

Print

Edit Forms

< Back to Records

Task Status

Errors (0)

Start

11-Dec-2024

Verify Complaint

13-Dec-2024

Investigate Complaint

17-Jan-2025

Active

Prompt Additional Regulatory Report

Void Additional Regulatory Report

Evaluate Regulatory

13-Dec-2024

Evaluate Product

External Evaluation

Ad Hoc

Verify Complaint Closure

Detail

Similar Cases

Product History Review

Investigation

Product Evaluation

Regulatory

Notes

Attachments

Cross References

Activity (6)

Complaint Information

Case Number

DP

CASE-2024-0020

Product

Demo Product

Aware Date

11-Dec-2024

Occurrence Date

11-Dec-2024

Lot Number

Lot Reference Number

UDI-DI Number

UDI-DI Issuing Entity

Basic UDI-DI

Basic UDI-DI Issuing Entity

Eudamed ID

Eudamed ID Issuing Entity

Eudamed-DI

Eudamed-DI Issuing Entity

Unit of Use UDI-DI

Unit of Use UDI-DI Issuing Entity

UDI/UPC Number

Expected Return Quantity

0

Decontamination Hold

No

Actual Quantity Returned

0

Actual Quantity Retained

Manufacturing Site

Manufactured Date

Expiration Date

Description Summary

Patient experienced adverse event possibly due to battery depletion

Failure Mode as Reported (0)

| Component | Failure Mode | IMDRF Medical Device Problem Codes (Annex A) | Root Cause | Root Cause Category | Primary Code |
|--------------------|--------------|--|------------|---------------------|--------------|
| No data to display | | | | | |

AI found similar cases with 80% or greater confidence. Click here to view in Detail panel.

Don't show again

★ CASE-2024-0020-1

Product: DP Country of Origin: UNITED STATES Reportable?: Yes Complaint Coordinator: TNEAL Reporting Site: NASH Aware Date: 11-Dec-2024 Complaint Aging Days: 254 Case Severity: Serious Injury Status: Inworks

Task ▾

Action ▾

Print ▾

Edit Forms ▾

Back to Records

Task Status Errors (0)

- Start
11-Dec-2024
- Verify Complaint
13-Dec-2024
- Investigate Complaint
17-Jan-2025
Active
- Prompt Additional Regulatory Report
- Void Additional Regulatory Report
- Evaluate Regulatory
13-Dec-2024
- Evaluate Product
- External Evaluation
- Ad Hoc
- Verify Complaint Closure

Detail Similar Cases Product History Review Investigation Product Evaluation Regulatory Notes Attachments Cross References Activity (6)

AI has identified previous case(s) that may be similar to the current based on their description summary.

| Record Number | Product | Occurrence Date | Description Summary | Confidence % | AI Creation Date |
|----------------|---------|-----------------|--|--------------|------------------|
| CASE-2024-0011 | DP | 11-Dec-2024 | Patient experienced adverse event possibly due to battery depletion | 100% | 11-Dec-2024 |
| CASE-2024-0019 | DP | 11-Dec-2024 | adverse event potentially due to battery discharge | 89.65% | 12-Dec-2024 |
| CASE-2024-0004 | DP | 10-Dec-2024 | adverse event potentially due to battery discharge. | 89.65% | 11-Dec-2024 |
| CASE-2024-0018 | DP | 11-Dec-2024 | Patient reported fainting due to premature battery depletion | 89.01% | 12-Dec-2024 |
| CASE-2024-0017 | DP | 11-Dec-2024 | Patient reported dizziness, possibly related to battery | 85.56% | 12-Dec-2024 |
| CASE-2024-0012 | DP | 11-Dec-2024 | Patient reported dizziness, possibly related to battery | 85.56% | 11-Dec-2024 |
| CASE-2024-0010 | DP | 11-Dec-2024 | Patient reported dizziness due to battery drain | 85.19% | 11-Dec-2024 |
| CASE-2024-0015 | DP | 11-Dec-2024 | Patient experienced lightheadedness and palpitations. It was found that the pacemaker's battery had depleted, leading to insufficient pacing. | 79.97% | 12-Dec-2024 |
| CASE-2024-0014 | DP | 11-Dec-2024 | Patient reported episodes of fatigue and chest discomfort. Investigation revealed a malfunction in the pacemaker's battery, causing it to intermittently stop working. | 77.87% | 12-Dec-2024 |
| CASE-2024-0013 | DP | 11-Dec-2024 | Patient experienced dizziness and shortness of breath due to the pacemaker's battery failure. The device was unable to maintain the required heart rate. | 76.35% | 12-Dec-2024 |
| CASE-2024-0016 | DP | 11-Dec-2024 | Patient suffered from weakness and irregular heartbeats due to a depleted pacemaker battery. The device failed to provide consistent pacing. | 74.44% | 12-Dec-2024 |

[← Back to Records](#)

Edit Forms ▼

Detail **Similar Cases** Product History Review Investigation Product Evaluation Regulatory Notes Attachments Cross References Activity (6)

| Record Number | Product | Occurrence Date | Description Summary | Confidence % | AI Creation Date |
|--------------------------------|---------|-----------------|--|--------------|------------------|
| CASE-2024-0011 | DP | 11-Dec-2024 | Patient experienced adverse event possibly due to battery depletion | 100% | 11-Dec-2024 |
| CASE-2024-0019 | DP | 11-Dec-2024 | adverse event potentially due to battery discharge | 89.65% | 12-Dec-2024 |
| CASE-2024-0004 | DP | 10-Dec-2024 | adverse event potentially due to battery discharge. | 89.65% | 11-Dec-2024 |
| CASE-2024-0018 | DP | 11-Dec-2024 | Patient reported fainting due to premature battery depletion | 89.01% | 12-Dec-2024 |
| CASE-2024-0017 | DP | 11-Dec-2024 | Patient reported dizziness, possibly related to battery | 85.56% | 12-Dec-2024 |
| CASE-2024-0012 | DP | 11-Dec-2024 | Patient reported dizziness, possibly related to battery | 85.56% | 11-Dec-2024 |
| CASE-2024-0010 | DP | 11-Dec-2024 | Patient reported dizziness due to battery drain | 85.19% | 11-Dec-2024 |
| CASE-2024-0015 | DP | 11-Dec-2024 | Patient experienced lightheadedness and palpitations. It was found that the pacemaker's battery had depleted, leading to insufficient pacing. | 79.97% | 12-Dec-2024 |
| CASE-2024-0014 | DP | 11-Dec-2024 | Patient reported episodes of fatigue and chest discomfort. Investigation revealed a malfunction in the pacemaker's battery, causing it to intermittently stop working. | 77.87% | 12-Dec-2024 |
| CASE-2024-0013 | DP | 11-Dec-2024 | Patient experienced dizziness and shortness of breath due to the pacemaker's battery failure. The device was unable to maintain the required heart rate. | 76.35% | 12-Dec-2024 |
| CASE-2024-0016 | DP | 11-Dec-2024 | Patient suffered from weakness and irregular heartbeats due to a depleted pacemaker battery. The device failed to provide consistent pacing. | 74.44% | 12-Dec-2024 |

- Dashboard
- Create
- Find
- Favorites
- Recent
- Calendar
-
- Admin
- System Setup
-

★ CASE-2024-0008-1

Investigate Complaint

Product: DP Country of Origin: UNITED STATES Reportable?: TBD Complaint Coordinator: TNEAL Reporting Site: NASH Aware Date: 10-Dec-2024 Complaint Aging Days: 255 Case Severity: Serious Injury

Status: Inworks

- Task ▾
- Action ▾
- Print ▾
- Attachment
- Edit Forms ▾

Task Status

Errors (0)

- Start
10-Dec-2024
- Verify Complaint
10-Dec-2024
- Plan Investigation
10-Dec-2024
- Investigate Complaint
14-Jan-2025 Active
- Approve Regulatory Evaluation
- Evaluate Product
- External Evaluation
- Ad Hoc
- Verify Complaint Closure
- Verify Regulatory

Investigate Complaint Task Information

Investigation Information

Product Information

Investigation Subtasks (4)

Note: To add or reopen subtasks, select 'Manage Investigation Subtasks' from the Action menu.

| Activity | Assigned To / Completed By | Due Date / Completed Date | Status |
|-------------------------------|----------------------------|---------------------------|-----------|
| Safety Investigation | TNEAL | 10-Dec-2024 | Completed |
| Manufacturing Investigation | TNEAL | 10-Dec-2024 | Completed |
| Quality Control Investigation | TNEAL | 10-Dec-2024 | Completed |
| Distribution Investigation | TNEAL | 10-Dec-2024 | Completed |

Note: Click 'Add' to record investigation results without routing a subtask. To add subtasks, select 'Manage Investigation Subtasks' from the Action menu.

Additional Results of Investigation

IMDRF Cause Investigation: Type of Investigation (Annex B)

Investigation Tools (0)

Failure Mode as Reported (0)

Root Cause Analysis

- Cancel
- Save
- Sign-off

- Detail
- Similar Cases
- Product History Review
- Investigation
- Product Evaluation
- Regulatory

Investigation Information

Investigation Subtasks

Safety Investigation (1 of 4)

Instructions

Result of Investigation

Assess the immediate and long-term risks to patients who may have been injected with the contaminated solution

Risk Assessment of Particulate Contamination

Scientific Literature Review indicate injected glass particles >50 microns can cause local tissue damage (granulomas, abscesses) and systemic complications (embolism, organ damage). Documented cases of similar incidents in the past have resulted in serious patient harm.

Patient Records Review

Medical records of 12 patients who received Drug Y from the implicated batch were reviewed. Two patients showed signs of local inflammation at the injection site, while one patient exhibited signs of phlebitis. There was found a correlation between inflammation cases with the presence of glass particles. No severe systemic events were reported, but early signs of tissue irritation were found.

Expert Consultation

Toxicologist and Vascular Specialist confirmed the risk of embolism if glass particles enter the bloodstream and advised immediate recall. Experts provided written recommendations for patient monitoring and recall procedures.

Immediate Actions

Notified healthcare providers of the issue, issued a recall of the affected batches, and instructed facilities to inspect vials before use. Recall Notice sent to all distributors and hospitals. A helpline was established for healthcare providers to report findings or concerns.

Completed By

Completed By Name

Completed Date

TNEAL

Todd Neal

10-Dec-2024

Attachments

★ CASE-2024-0008-1

Investigate Complaint

Product: DP Country of Origin: UNITED STATES Reportable?: TBD Complaint Coordinator: TNEAL Reporting Site: NASH Aware Date: 10-Dec-2024 Complaint Aging Days: 255 Case Severity: Serious Injury

Status: Inworks

Task

Action

Print

Attachment

Edit Forms

Task Status Errors (0)

Start

10-Dec-2024

Verify Complaint

10-Dec-2024

Plan Investigation

10-Dec-2024

Investigate Complaint

14-Jan-2025

Active

Approve Regulatory Evaluation

Evaluate Product

External Evaluation

Ad Hoc

Verify Complaint Closure

Verify Regulatory

Investigate Complaint Task Information

Investigation Information

Product Information

Investigation Subtasks (4)

Note: To add or reopen subtasks, select 'Manage Investigation Subtasks' from the Action menu.

| Activity | Assigned To / Completed By | Due Date / Completed Date | Status |
|-------------------------------|----------------------------|---------------------------|-----------|
| Safety Investigation | TNEAL | 10-Dec-2024 | Completed |
| Manufacturing Investigation | TNEAL | 10-Dec-2024 | Completed |
| Quality Control Investigation | TNEAL | 10-Dec-2024 | Completed |
| Distribution Investigation | TNEAL | 10-Dec-2024 | Completed |

Note: Click 'Add' to record investigation results without routing a subtask. To add subtasks, select 'Manage Investigation Subtasks' from the Action menu.

Additional Results of Investigation

IMDRF Cause Investigation: Type of Investigation (Annex B)

Investigation Tools (0)

Failure Mode as Reported (0)

Root Cause Analysis

Cancel

Save

Sign-off

Manufacturing Investigation (2 of 4)

Instructions

Result of Investigation

Determine how glass particles were introduced into vials during manufacturing process

Equipment Inspection

Inspected vial-filling and capping machines using high-resolution cameras and borescopes. It was discovered that minor wear on filling nozzles could have generated glass particles. Nozzles were replaced.

Vial Integrity Testing

Stress tests were conducted on empty vials to assess durability. Vials from the implicated batch showed a higher breakage rate under stress. Therefore, vial breakage was confirmed during the filling process, leading to glass particle contamination.

Raw Material Review

Certificates of Analysis (COA) from the glass vial supplier and inspected incoming vials were reviewed. It was identified that a defective batch of vials from the supplier did not meet pharmaceutical-grade standards. Future orders were suspended pending corrective actions and supplier audit.

Immediate Actions

Replaced the worn filling nozzles and quarantined all vials from the defective supplier lot: 12,000 vials from the defective lot were quarantined. Established a preventive maintenance program for all vial-filling and sealing equipment, including regular inspections and calibration. Equipment inspection records are now reviewed monthly by the quality team.

Completed By

Completed By Name

Completed Date

TNEAL

Todd Neal

10-Dec-2024

Attachments

Quality Control Investigation (3 of 4)

Instructions

Result of Investigation

Confirm the presence of foreign particles in the drug product and assess any failures in QC protocols

Visual and Microscopic Inspection

Conducted 100% visual inspection of the implicated vials under controlled lighting and magnification. Glass particles were confirmed in 3% of the inspected vials from the implicated batch.

- Dashboard
- Create
- Find
- Favorites
- Recent
- Calendar
-
- Admin
- System Setup

★ CASE-2024-0008-1

Investigate Complaint

Product: DP Country of Origin: UNITED STATES Reportable?: TBD Complaint Coordinator: TNEAL Reporting Site: NASH Aware Date: 10-Dec-2024 Complaint Aging Days: 255 Case Severity: Serious Injury

Status: Inworks

- Task
- Action
- Print
- Attachment
- Edit Forms

Task Status Errors (0)

Start

10-Dec-2024

Verify Complaint

10-Dec-2024

Plan Investigation

10-Dec-2024

Investigate Complaint

14-Jan-2025

Active

Approve Regulatory Evaluation

Evaluate Product

External Evaluation

Ad Hoc

Verify Complaint Closure

Verify Regulatory

Investigate Complaint Task Information

Investigation Information

Product Information

Investigation Subtasks (4)

Note: To add or reopen subtasks, select 'Manage Investigation Subtasks' from the Action menu.

| Activity | Assigned To / Completed By | Due Date / Completed Date | Status |
|-------------------------------|----------------------------|---------------------------|-----------|
| Safety Investigation | TNEAL | 10-Dec-2024 | Completed |
| Manufacturing Investigation | TNEAL | 10-Dec-2024 | Completed |
| Quality Control Investigation | TNEAL | 10-Dec-2024 | Completed |
| Distribution Investigation | TNEAL | 10-Dec-2024 | Completed |

Note: Click 'Add' to record investigation results without routing a subtask. To add subtasks, select 'Manage Investigation Subtasks' from the Action menu.

Additional Results of Investigation

IMDRF Cause Investigation: Type of Investigation (Annex B)

Investigation Tools (0)

Failure Mode as Reported (0)

Root Cause Analysis

Cancel

Save

Sign-off

Quality Control Investigation (3 of 4)

Instructions

Confirm the presence of foreign particles in the drug product and assess any failures in QC protocols

Visual and Microscopic Inspection

Conducted 100% visual inspection of the implicated vials under controlled lighting and magnification. Glass particles were confirmed in 3% of the inspected vials from the implicated batch.

Chemical Analysis

Performed an energy-dispersive X-ray spectroscopy (EDS) to confirm the composition of the particles. EDS confirmed the particles were glass, matching the composition of the defective vials.

Particle Count Test

Conducted a particle count test according to USP <788> standards. The implicated batch exceeded the acceptable limits for particulate matter, confirming a QC failure.

Immediate Actions

Quarantined the affected batch. New QC SOPs were issued, and QC personnel underwent training on inspection techniques. Implemented an automated particle detection system to supplement manual inspections.

Completed By

Completed By Name

Completed Date

TNEAL

Todd Neal

10-Dec-2024

Attachments

Distribution Investigation (4 of 4)

Instructions

Determine the extent of distribution of the implicated batches and assess the need for a recall

Batch Tracking

Used the batch number to track distribution to wholesalers, pharmacies, and hospitals. Implicated batches were distributed to 15 hospitals and 20 pharmacies across 5 states.

Inventory Check

Contacted distributors and healthcare facilities to assess if any remaining stock was available. 75% of the implicated batch was recovered. The remaining 25% had already been administered, prompting a patient notification process.

- Dashboard
- Create
- Find
- Favorites
- Recent
- Calendar
-
- Admin
- System Setup
-

★ CASE-2024-0008-1

Investigate Complaint

Product: DP Country of Origin: UNITED STATES Reportable?: TBD Complaint Coordinator: TNEAL Reporting Site: NASH Aware Date: 10-Dec-2024 Complaint Aging Days: 255 Case Severity: Serious Injury

Status: Inworks


- Task
- Action
-  Print
-  Attachment
-  Edit Forms

Task Status

Errors (0)


Start

10-Dec-2024




Verify Complaint

10-Dec-2024



Plan Investigation

10-Dec-2024



Investigate Complaint

14-Jan-2025

Active

Approve Regulatory Evaluation

Evaluate Product

External Evaluation

Ad Hoc

Verify Complaint Closure

Verify Regulatory

Investigate Complaint Task Information

Investigation Information

Product Information

Investigation Subtasks (4)

Note: To add or reopen subtasks, select 'Manage Investigation Subtasks' from the Action menu.

| Activity | Assigned To / Completed By | Due Date / Completed Date | Status | |
|-------------------------------|----------------------------|---------------------------|-----------|---|
| Safety Investigation | TNEAL | 10-Dec-2024 | Completed | > |
| Manufacturing Investigation | TNEAL | 10-Dec-2024 | Completed | > |
| Quality Control Investigation | TNEAL | 10-Dec-2024 | Completed | > |
| Distribution Investigation | TNEAL | 10-Dec-2024 | Completed | > |

Note: Click 'Add' to record investigation results without routing a subtask. To add subtasks, select 'Manage Investigation Subtasks' from the Action menu.

Additional Results of Investigation

IMDRF Cause Investigation: Type of Investigation (Annex B)

Investigation Tools (0)

Failure Mode as Reported (0)

Root Cause Analysis

Cancel

Save

Sign-off

Distribution Investigation (4 of 4)

Instructions

Determine the extent of distribution of the implicated batches and assess the need for a recall

Batch Tracking

Used the batch number to track distribution to wholesalers, pharmacies, and hospitals. Implicated batches were distributed to 15 hospitals and 20 pharmacies across 5 states.

Inventory Check

Contacted distributors and healthcare facilities to assess if any remaining stock was available. 75% of the implicated batch was recovered. The remaining 25% had already been administered, prompting a patient notification process.

Recall Risk Assessment

Conducted a risk-benefit analysis to determine the necessity of a recall. The recall was classified as a Class II recall (FDA), and the necessary notifications were made to regulatory authorities. Recall process was completed within 2 weeks. A final recall report was submitted to the FDA.

Completed By

TNEAL

Completed By Name

Todd Neal

Completed Date

10-Dec-2024

Attachments

Additional Results of Investigation

IMDRF Cause Investigation: Type of Investigation (Annex B)

Investigation Tools (0)

Failure Mode as Reported (0)

IMDRF Cause Investigation: Investigation Findings (Annex C)

IMDRF Component Codes (Annex G)

Dashboard

Create

Find

Favorites

Recent

Calendar

Admin

System Setup

★ CASE-2024-0008-1

Investigate Complaint

Product: DP Country of Origin: UNITED STATES Reportable?: TBD Complaint Coordinator: TNEAL Reporting Site: NASH Aware Date: 10-Dec-2024 Complaint Aging Days: 255 Case Severity: Serious Injury

Status: Inworks

Task

Action

Print

Attachment

Edit Forms

Task Status Errors (0)

Start

10-Dec-2024

Verify Complaint

10-Dec-2024

Plan Investigation

10-Dec-2024

Investigate Complaint

14-Jan-2025

Active

Approve Regulatory Evaluation

Evaluate Product

External Evaluation

Ad Hoc

Verify Complaint Closure

Verify Regulatory

IMDRF Component Codes (Annex G)

CAPA Escalation

Action Plans (0)

Conclusion Code (1)

Investigation Summary

Confirm Complaint?

Yes No

AI can draft an investigation summary based on Case Description and Investigation Results.

Generate Summary

Attachments (0)

Cancel

Save

Sign-off

Detail Similar Cases Product History Review Investigation Product Evaluation Regulatory

Investigation Information

Investigation Subtasks

Safety Investigation (1 of 4)

Instructions

Result of Investigation

Assess the immediate and long-term risks to patients who may have been injected with the contaminated solution

Risk Assessment of Particulate Contamination
Scientific Literature Review indicate injected glass particles >50 microns can cause local tissue damage (granulomas, abscesses) and systemic complications (embolism, organ damage). Documented cases of similar incidents in the past have resulted in serious patient harm.

Patient Records Review
Medical records of 12 patients who received Drug Y from the implicated batch were reviewed. Two patients showed signs of local inflammation at the injection site, while one patient exhibited signs of phlebitis. There was found a correlation between inflammation cases with the presence of glass particles. No severe systemic events were reported, but early signs of tissue irritation were found.

Expert Consultation
Toxicologist and Vascular Specialist confirmed the risk of embolism if glass particles enter the bloodstream and advised immediate recall. Experts provided written recommendations for patient monitoring and recall procedures.

Immediate Actions
Notified healthcare providers of the issue, issued a recall of the affected batches, and instructed facilities to inspect vials before use. Recall Notice sent to all distributors and hospitals. A helpline was established for healthcare providers to report findings or concerns.

Completed By

Completed By Name

Completed Date

TNEAL

Todd Neal

10-Dec-2024

Attachments

- Dashboard
- Create
- Find
- Favorites
- Recent
- Calendar
-
- Admin
- System Setup
-

★ CASE-2024-0008-1

Investigate Complaint

Product: DP Country of Origin: UNITED STATES Reportable?: TBD Complaint Coordinator: TNEAL Reporting Site: NASH Aware Date: 10-Dec-2024 Complaint Aging Days: 255 Case Severity: Serious Injury

Status: Inworks

- Task
- Action
- Print
- Attachment
- Edit Forms

Task Status Errors (0)

Start

10-Dec-2024

Verify Complaint

10-Dec-2024

Plan Investigation

10-Dec-2024

Investigate Complaint

14-Jan-2025

Active

Approve Regulatory Evaluation

Evaluate Product

External Evaluation

Ad Hoc

Verify Complaint Closure

Verify Regulatory

IMDRF Component Codes (Annex G)

CAPA Escalation

Action Plans (0)

Conclusion Code (1)

Investigation Summary

Confirm Complaint?

Yes No

AI can draft an investigation summary based on Case Description and Investigation Results.

Generate Summary

AI drafted on: 22-Aug-2025 08:06

Probable Root Cause: The investigation identified that the presence of glass particles in vials of Drug Y was due to minor wear on the filling nozzles and defective vials from a supplier that did not meet pharmaceutical-grade standards. Equipment inspection revealed that the worn nozzles could generate glass particles, and stress tests confirmed vial breakage during the filling process.

Impact: The presence of glass particles posed a risk of local tissue damage and systemic complications, such as embolism and organ damage. Medical records review showed signs of local inflammation in two patients and phlebitis in one patient, correlating with the presence of glass particles. No severe systemic events were reported, but early signs of tissue irritation were found.

Conclusion: Immediate actions were taken, including notifying healthcare providers, issuing a recall of affected batches, and implementing preventive measures. Equipment was repaired, and quality control processes were enhanced. A Class II recall was completed, and a final report was submitted to the FDA.

Cancel

Save

Sign-off

Detail

Similar Cases

Product History Review

Investigation

Product Evaluation

Regulatory

Investigation Information

Investigation Subtasks

Safety Investigation (1 of 4)

Instructions

Result of Investigation

Assess the immediate and long-term risks to patients who may have been injected with the contaminated solution

Risk Assessment of Particulate Contamination

Scientific Literature Review indicate injected glass particles >50 microns can cause local tissue damage (granulomas, abscesses) and systemic complications (embolism, organ damage). Documented cases of similar incidents in the past have resulted in serious patient harm.

Patient Records Review

Medical records of 12 patients who received Drug Y from the implicated batch were reviewed. Two patients showed signs of local inflammation at the injection site, while one patient exhibited signs of phlebitis. There was found a correlation between inflammation cases with the presence of glass particles. No severe systemic events were reported, but early signs of tissue irritation were found.

Expert Consultation

Toxicologist and Vascular Specialist confirmed the risk of embolism if glass particles enter the bloodstream and advised immediate recall. Experts provided written recommendations for patient monitoring and recall procedures.

Immediate Actions

Notified healthcare providers of the issue, issued a recall of the affected batches, and instructed facilities to inspect vials before use. Recall Notice sent to all distributors and hospitals. A helpline was established for healthcare providers to report findings or concerns.

Completed By

Completed By Name

Completed Date

TNEAL

Todd Neal

10-Dec-2024

Attachments

Complaint Setup

Show Form Delta Changes Edit Mode

Complaint Setup

- Apply to All Product Types
- Apply to All Reporting Sites
- Apply to All Case Sources

Team *

Complaint Team - NASH

Complaint Team - NASH

Investigation Summary - AI Prompt

Generate a summary including all of the following sections:
- Probable Root Cause
- Impact
- Conclusion

Policy

Due Date (# of business days)

Cancel

Delete

Submit



Thank you!

Please contact us for more information